

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 07/29/2020

OPERATOR NAME:  
 RICHARD J ARNOLD  
 Permit #: 33860  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:15
INTERNAL STANDARD	VERIFIED	01:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:15
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:16
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:18
BLANK TEST	0.00	01:19

