

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 08/10/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:19
INTERNAL STANDARD	VERIFIED	06:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:19
BLANK TEST	0.00	06:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:20
BLANK TEST	0.00	06:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:21
BLANK TEST	0.00	06:22

