

SUPERVISOR TEST

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 08/12/2020

OPERATOR NAME:
 BRAD M LYON
 Permit #: 44978
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:56
INTERNAL STANDARD	VERIFIED	07:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:56
BLANK TEST	0.00	07:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:59
BLANK TEST	0.00	08:00

