

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 08/02/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:19
INTERNAL STANDARD	VERIFIED	09:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:20
BLANK TEST	0.00	09:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:21
BLANK TEST	0.00	09:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:22
BLANK TEST	0.00	09:23

