

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 07/27/2020

OPERATOR NAME:
ERIC C WRIGHT
Permit #: 49047
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:54
INTERNAL STANDARD	VERIFIED	13:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:55
BLANK TEST	0.00	13:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:56
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58

