

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/03/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:12
INTERNAL STANDARD	VERIFIED	22:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:12
BLANK TEST	0.00	22:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:15
BLANK TEST	0.00	22:16

