

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/26/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:56
INTERNAL STANDARD	VERIFIED	18:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:57
BLANK TEST	0.00	18:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:58
BLANK TEST	0.00	18:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00

