

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 09/20/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:02
INTERNAL STANDARD	VERIFIED	02:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:04
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:05
BLANK TEST	0.00	02:06

