

SOLUTION CHANGE

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dnt: 126706

Date: 09/01/2020

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:40
INTERNAL STANDARD	VERIFIED	12:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:41
BLANK TEST	0.00	12:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:42
BLANK TEST	0.00	12:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:43
BLANK TEST	0.00	12:44

