

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 08/25/2020

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:05
INTERNAL STANDARD	VERIFIED	07:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:06
BLANK TEST	0.00	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:07
BLANK TEST	0.00	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:08
BLANK TEST	0.00	07:09

