

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dnt: 126706

Date: 09/08/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:25
INTERNAL STANDARD	VERIFIED	13:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:25
BLANK TEST	0.00	13:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:26
BLANK TEST	0.00	13:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:27
BLANK TEST	0.00	13:28

