

SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 11/02/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:17
INTERNAL STANDARD	VERIFIED	10:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:19
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:20
BLANK TEST	0.00	10:21

