

SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 11/08/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:09
INTERNAL STANDARD	VERIFIED	18:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:10
BLANK TEST	0.00	18:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:11
BLANK TEST	0.00	18:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:12
BLANK TEST	0.00	18:13

