

SOLUTION CHANGE

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 09/01/2020

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:03
INTERNAL STANDARD	VERIFIED	20:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:05
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08

