

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 09/22/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:06
INTERNAL STANDARD	VERIFIED	19:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:07
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10

