

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 09/28/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:58
INTERNAL STANDARD	VERIFIED	14:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:59
BLANK TEST	0.00	15:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:00
BLANK TEST	0.00	15:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:02
BLANK TEST	0.00	15:03

