

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 10/12/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:17
INTERNAL STANDARD	VERIFIED	19:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:18
BLANK TEST	0.00	19:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:19
BLANK TEST	0.00	19:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:20
BLANK TEST	0.00	19:21

