

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/18/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 23523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:23
INTERNAL STANDARD	VERIFIED	19:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:24
BLANK TEST	0.00	19:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:25
BLANK TEST	0.00	19:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:27
BLANK TEST	0.00	19:28

