

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 11/14/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:32
INTERNAL STANDARD	VERIFIED	23:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:32
BLANK TEST	0.00	23:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:33
BLANK TEST	0.00	23:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:34
BLANK TEST	0.00	23:35

