

SOLUTION CHANGE

GRANVILLE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 113306

Date: 11/01/2020

OPERATOR NAME:
 RYAN S PEDONE
 Permit #: 35878
 Operator Agency/Dept: GRANVILLE VILLAGE PD
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:19
INTERNAL STANDARD	VERIFIED	23:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:19
BLANK TEST	0.00	23:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:20
BLANK TEST	0.00	23:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:22
BLANK TEST	0.00	23:22

