

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/18/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:07
INTERNAL STANDARD	VERIFIED	02:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:09
BLANK TEST	0.00	02:10

