

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 10/24/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:56
INTERNAL STANDARD	VERIFIED	19:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:56
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:58
BLANK TEST	0.00	19:59

