

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 11/15/2020

OPERATOR NAME:
BRAD M LYON
Permit #: 44978
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:36
INTERNAL STANDARD	VERIFIED	08:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:37
BLANK TEST	0.00	08:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:39
BLANK TEST	0.00	08:40

