

SOLUTION CHANGE

-----  
HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408  
-----

Date: 11/04/2020

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:22
INTERNAL STANDARD	VERIFIED	00:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:24
BLANK TEST	0.00	00:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:25
BLANK TEST	0.00	00:26

