

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 09/22/2020

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:55
INTERNAL STANDARD	VERIFIED	20:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:55
BLANK TEST	0.00	20:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:56
BLANK TEST	0.00	20:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:58
BLANK TEST	0.00	20:58

