

SUPERVISOR TEST

-----  
 WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806  
 -----

Date: 10/31/2020

OPERATOR NAME:  
 BRYAN J GRECO  
 Permit #: 44077  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:19
INTERNAL STANDARD	VERIFIED	06:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:20
BLANK TEST	0.00	06:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:21
BLANK TEST	0.00	06:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:22
BLANK TEST	0.00	06:23

