

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 10/24/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20340

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 12:07 |
| INTERNAL STANDARD | VERIFIED | 12:07 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 12:08 |
| BLANK TEST | 0.00 | 12:09 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 12:09 |
| BLANK TEST | 0.00 | 12:10 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 12:10 |
| BLANK TEST | 0.00 | 12:11 |

