

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 09/12/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:27
INTERNAL STANDARD	VERIFIED	07:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:29
BLANK TEST	0.00	07:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:30
BLANK TEST	0.00	07:31

