

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 08/23/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:58
INTERNAL STANDARD	VERIFIED	06:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:58
BLANK TEST	0.00	06:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:59
BLANK TEST	0.00	07:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:00
BLANK TEST	0.00	07:01

