

# SOLUTION CHANGE

-----  
CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706  
-----

Date: 11/30/2020

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:38
INTERNAL STANDARD	VERIFIED	10:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:39
BLANK TEST	0.00	10:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:40
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:41
BLANK TEST	0.00	10:42

