

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 12/26/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:17
INTERNAL STANDARD	VERIFIED	06:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:18
BLANK TEST	0.00	06:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:19
BLANK TEST	0.00	06:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:20
BLANK TEST	0.00	06:21

