

WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 12/14/2020

OPERATOR NAME:
 BRYAN J GRECO
 Permit #: 44077
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:52
INTERNAL STANDARD	VERIFIED	06:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:53
BLANK TEST	0.00	06:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:54
BLANK TEST	0.00	06:55

