

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 11/28/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:18
INTERNAL STANDARD	VERIFIED	06:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:19
BLANK TEST	0.00	06:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:20
BLANK TEST	0.00	06:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:21
BLANK TEST	0.00	06:22

