

SUPERVISOR TEST

WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 08/18/2020

OPERATOR NAME:
 JASON M VANDENBURGH
 Permit #: 37635
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 20240

Nominal ~ 0.10 g/210L

BLANK TEST	0.00	06:05
INTERNAL STANDARD	VERIFIED	06:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:05
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:07
BLANK TEST	0.00	06:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:08
BLANK TEST	0.00	06:09

