

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 11/21/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:49
INTERNAL STANDARD	VERIFIED	06:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:50
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:52
BLANK TEST	0.00	06:52

