

SUPERVISOR TEST

-----  
 FORT EDWARD  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 115708  
 -----

Date: 11/22/2020

OPERATOR NAME:  
 SGT DEAN E WATKINS  
 Permit #: 39523  
 Operator Agency/Dept: FORT EDWARD VILL. PD  
 Reference Standard #: 20340

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 21:27 |
| INTERNAL STANDARD       | VERIFIED | 21:27 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 21:27 |
| BLANK TEST              | 0.00     | 21:28 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 21:29 |
| BLANK TEST              | 0.00     | 21:29 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 21:30 |
| BLANK TEST              | 0.00     | 21:31 |

