

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 12/20/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:53
INTERNAL STANDARD	VERIFIED	20:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:54
BLANK TEST	0.00	20:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:55
BLANK TEST	0.00	20:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:57
BLANK TEST	0.00	20:57

