

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/08/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:04
INTERNAL STANDARD	VERIFIED	21:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:04
BLANK TEST	0.00	21:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:05
BLANK TEST	0.00	21:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:07
BLANK TEST	0.00	21:08

