

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 11/15/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:45
INTERNAL STANDARD	VERIFIED	20:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:46
BLANK TEST	0.00	20:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:47
BLANK TEST	0.00	20:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:48
BLANK TEST	0.00	20:49

