

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/19/2020

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:41
INTERNAL STANDARD	VERIFIED	23:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:43
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45

(Handwritten signature) #691

