

SUPERVISOR TEST

WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406

Date: 09/21/2020

OPERATOR NAME:  
 WILLIAM M GREENE  
 Permit #: 48760  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:40
INTERNAL STANDARD	VERIFIED	16:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:41
BLANK TEST	0.00	16:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:42
BLANK TEST	0.00	16:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:43
BLANK TEST	0.00	16:44

*WS*  
*#647*

