

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 11/01/2020

OPERATOR NAME:  
TRAVIS A EARL  
Permit #: 33337  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:14
INTERNAL STANDARD	VERIFIED	12:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:14
BLANK TEST	0.00	12:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:16
BLANK TEST	0.00	12:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:17
BLANK TEST	0.00	12:18

*T. Earl #680*

