

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/09/2020

OPERATOR NAME:
NICHOLAS D HOFFMAN
Permit #: 49232
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:12
INTERNAL STANDARD	VERIFIED	21:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:13
BLANK TEST	0.00	21:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:14
BLANK TEST	0.00	21:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:15
BLANK TEST	0.00	21:16

Nicholas D Hoffman
641

