

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/21/2020

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:48
INTERNAL STANDARD	VERIFIED	06:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52

CS
#603

