

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/24/2020

OPERATOR NAME:
TRAVIS A EARL
Permit #: 33337
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:43
INTERNAL STANDARD	VERIFIED	13:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:44
BLANK TEST	0.00	13:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:45
BLANK TEST	0.00	13:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:46
BLANK TEST	0.00	13:47

J.A. Earl
#680

