

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 10/05/2020

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:46
INTERNAL STANDARD	VERIFIED	04:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:47
BLANK TEST	0.00	04:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:48
BLANK TEST	0.00	04:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:49
BLANK TEST	0.00	04:50

