

SUPERVISOR TEST

-----  
 WASHINGTON CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100284  
 -----

Date: 10/12/2020

OPERATOR NAME:  
 MATTHEW K JACKSON  
 Permit #: 40078  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:02
INTERNAL STANDARD	VERIFIED	13:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:03
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:04
BLANK TEST	0.00	13:05

