

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 10/26/2020

OPERATOR NAME:
 MATTHEW K JACKSON
 Permit #: 40078
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:10
INTERNAL STANDARD	VERIFIED	11:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:11
BLANK TEST	0.00	11:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:12
BLANK TEST	0.00	11:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:13
BLANK TEST	0.00	11:14

