

SUPERVISOR TEST

-----  
WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284  
-----

Date: 12/21/2020

OPERATOR NAME:  
ERIC C WRIGHT  
Permit #: 49047  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20430

Nominal = 0.10 g/210L

RFI Detected

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 12/22/2020

OPERATOR NAME:  
MATTHEW K JACKSON  
Permit #: 40078  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:03
INTERNAL STANDARD	VERIFIED	00:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07

