

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/28/2020

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:20
INTERNAL STANDARD	VERIFIED	18:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:22
BLANK TEST	0.00	18:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:23
BLANK TEST	0.00	18:24

